

Smith & Nephew Medical Ltd.
101 Hessle Road
Hull, HU3 2BN
England

T + 44 (0) 1482 225181
F + 44 (0) 1482 328326
www.smith-nephew.com



EUROPEAN DECLARATION OF CONFORMITY

This Declaration confirms that the product listed below meets the Essential Requirements set out in Annex I of the Council Directive 93/42/EEC (as amended).

Manufacturer's Name: Smith and Nephew Medical Limited
Business Address: 101 Hessle Road,
Hull
HU3 2BN
United Kingdom

Medical Devices: ALLEVYN Life
Classification: Class IIb Sterile
GMDN Code and Term: 44970 Dressing, highly absorbent
Scope of Application: All batches supplied to which the Declaration of Conformity Procedure has been applied.

Declaration : Conformity of the product has been assessed in accordance with Annex II of the Directive. A dossier of technical documentation, as required by the Directive is available. The product listed is designed, manufactured and tested in accordance with the information set out in the dossier.

Verification Certificate(s): EC Certificate No. CE 00356 Full Quality Assurance.
Notified Body No. 0086 (British Standards Institute)
British Standards Institute. Certificate No. MD 76718 Quality Management System (BS EN ISO 13485)
British Standards Institute. Certificate No. FM 24676 Quality Management System (BS EN ISO 9001)

Standards Applied:

BS EN 556-1:2001/AC:2006	BS EN ISO 780:2015
BS EN 980: 2008	BS EN 1041: 2008
BS EN ISO 9001: 2008	BS EN ISO 10993-1:2009 /AC: 2010
BS EN ISO 10993-2:2006	BS EN ISO 10993-3:2014
BS EN ISO 10993-5:2009	BS EN ISO 10993-7:2008/ AC:2009
BS EN ISO 10993-10:2013	BS EN ISO 10993-12:2012
BS EN ISO 10993-18: 2009	BS EN ISO 11135:2014
BS EN ISO 11607-1: 2009	BS EN ISO 11607-2:2006
BS EN ISO 13485: 2012	BS EN 13726-1:2002/AC:2003
BS EN 13726-2: 2002	BS EN 13726-3: 2003
BS EN 13726-4:2003	BS EN ISO 14644 -1:2016
BS EN ISO 14971: 2012	BS EN ISO 15223-1: 2012
ISTA 2A: 2011	

Authorised Signatory :

Name : Susan McLoughlin
Position : Director of Regulatory Affairs

Signed :

Dated :

23/08/2016

Certificate Reference

HU/062 issue17